

	सामग्री प्रबंधन अनुभाग भारतीय प्रौद्योगिकी संस्थान, रुडकी रुडकी-247667, हरिद्वार, उत्तराखण्ड, भारत Phone-(O) 01332-28-4293, 4693 E-mail: mmiitr@iitr.ac.in	Material Management Section Indian Institute of Technology Roorkee-247667( Haridwar) (Uttarakhand) (India) Phone-(O) 01332-28-4293, 4693 E-mail: mmiitr@iitr.ac.in	
	<b>GSTIN-05AAALI0033R42Z</b>	E-mail: mmiitr@iitr.ac.in	
	PAN-AAALI0033R		

**निविदा प्रलेख / Tender Document**

**No. 2025260425/MM-13/IITR/2025-26/Floor Mounted Powder X-ra/CHEMISTRY/448**

Date of Uploading of Tender -<https://eprocure.gov.in/eprocure/app>

**12-Dec-2025**

Bids under two bid systems (Technical and Commercial) are invited Online at CPP Portal (<https://eprocure.gov.in/eprocure/app>) from the reputed manufacturer/authorized supplier/dealer for the following items. **Manual bids will not be accepted.**

वस्तु की तकनीकी विशिष्टताएँ / Technical Specifications of item:

Sr. No	Name of item / Specifications	QTY	UOM
<b>1</b>	<b>Floor Mounted Powder X-ray Diffractometer</b>	<b>1</b>	<b>Set</b>
<b>(For Detailed Technical Specification Plz refer Annexure-V)</b>			
नोट-1: कृपया वस्तु का मूल्य एवं कर अलग-अलग दर्शाए।			
Note-1: Please quote the Rate & Taxes of the item separately.			
Note-2: E-way bill is to be issued by the Venders/Transporter			
Note-3: The HSN / SAC Code of the item must be mentioned.			
For Reasonability of Rates, The firm must enclose the last two supply orders for the same item with Financial Bid (Annexure-III) only.			
<b>Critical Date Sheet</b>			
1	Published Date	<b>12-Dec-2025</b>	<b>(18:30)</b>
2	Bid Document Download / Sale Start Date	<b>12-Dec-2025</b>	<b>(18:40)</b>
3	Seek Clarification Start Date	<b>12-Dec-2025</b>	<b>(18:50)</b>
4	Seek Clarification End Date	<b>17-Dec-2025</b>	<b>(15:00)</b>
5	Pre Bid Meeting Date	<b>Not Required</b>	<b>Not Required</b>
6	Bid Submission Start Date	<b>18-Dec-2025</b>	<b>(16:00)</b>
7	Bid Submission End Date	<b>2-Jan-2026</b>	<b>(15:00)</b>
8	Bid Opening Date	<b>2-Jan-2026</b>	<b>(15:30)</b>

**Chapter-1 : Instruction to bidders**

- A. The Tender should be enclosed with proper certifications like Agency Certification, Authorization certificate and/or Proprietary Certificate, as the case may be, in support of your offer.
- B. बोली का जमा किया जाना / **Submission of Bids:**  
**The bids should be submitted online in two parts –**
  - 1 (i) The Technical Bid (Annexure-I, II, IV & V) and Tender Acceptance Letter  
(ii) The Financial Bid (Annexure-III) & Price Bid in .XLS Format  
**Rates in .xls format will only be considered for Financial evaluation & further processing.**
  - 2 The bids should be on official pad preferably with GST Number of the firm

- 3 The quotations should bear full details and where possible may be duly supported with catalogues, pamphlets, literature, samples of the item/items as the case may be for comparing the quality and rates of the item(s).
- 4 The transportation cost, insurance charge etc., if any, percentage/ rate of GST or all other taxes and duties should be clearly mentioned.
- 5 The bids / rates / tenders should remain valid for a minimum period of 90 days from the date of financial opening.
- 6 The rates shall not be subject to escalation of any nature.
- 7 The rates quoted should be applicable to educational institutions and any cost advantage received in lieu thereof should be passed on the Institution.
- 8 While quoting / sending rates, the firm shall give an undertaking as per Annexure-II.

**C. विधि निक्षेप/ Earnest Money Deposit (EMD):**

**In lieu of Bid Security/Earnest Money Deposit (EMD), declaration as per Appendix-A should be provided.**

**D. निष्पादन बैंक प्रत्याभूति/ Performance Bank Guarantee (PBG):**

**In case the total order value of items becomes Rs.25.00 Lakhs or above, 5% of the total order value in the form of Bank Guarantee or FDR pledged in favor of "The Registrar, IIT Roorkee" is required at the time of installation of the equipment which should remain valid at least during the warranty period plus sixty days. No interest shall be claimed by the bidder on the PBG submitted.**

**E. बोलियों का खोला जाना/ Opening of Bids:**

- 1 The Technical Bids will be opened online at CPPP Portal (<https://eprocure.gov.in/eprocure/app>)-Plz refer **Critical Date Sheet**
- 2 The Financial bids of only the Technically Qualified bidders will be opened. The date and time for opening the Financial Bid will be communicated to the respective firm later by the system (CPP Portal (<https://eprocure.gov.in/eprocure/app>))
- 3 Representatives of the firms(s) may be present at the time of openings, if they so desire.

**Chapter-2 : Conditions of Contract**

**F. ठेका देना/ Award of Contract:**

The final selection of the bidder for the award of the contract will be made on the basis of the lowest commercial bid amongst the Technically Qualified Bidders.

**Prices of optional item will not be considered to identify lowest price bid.**

**G. पूर्व-आपूर्ति निरीक्षण/ Pre-supply Inspection:**

Authorized representative of the Institute shall make the final inspection before supply of the item at site of the firm, if required.

**H. वस्तु की आपूर्ति/ Supply of Item:**

The whole supply as per order shall have to be completed within the time mentioned in the order failing which the I.I.T. Roorkee shall have the right to accept or reject any quantity of items ordered. The firm will have to arrange for supply of the material in good condition.

**I. भुगतान की शर्तें/ Payment Terms:**

**A. Payment in INR for supply of goods:**

Where installation is not required: 100% payment will be made after satisfactory delivery of material duly certified by the HOD/P.I.

**OR**

Where Installation is required :100% payment will be made after satisfactory delivery and installation of the material duly certified by the HOD/P.I.

**OR**

Upto 80% payment on receipt and acceptance of goods/service by the indenter and the balance on successful installation and commissioning by the supplier and acceptance by the indenter.

**B. Payment in INR for AMC/Service:**

Quarterly/Halfyearly after satisfactory completion of work/service duly certified by the H.O.D./P.I.

**J. दंड/Penalty:**

A penalty of 0.5% of the total order value per week shall be levied for the delay subject to a maximum of 5% of the total order value.

In case of the non-supply of the item/service after the acceptance of the work/supply order, EMD & PBG (if submitted) will be forfeited by the Institute. Also, the firm will be debarred from any future bidding process of IIT Roorkee for a period of a minimum two year.

**K. अप्रत्याशित घटना/Force Majeure:**

For the purpose of and within the scope of contract by way of indication and not of limitation, the term “Force Majeure” shall mean acts of nature, strikes, Lockouts, or other industrial disturbances, act of public/enemy, wars, blockades, insurrection, riots, epidemics, landslides, earthquakes, storms lightening, flood, washouts, civil disturbances, explosion and any other similar even not within the control of either party on which by exercise of due care and diligence neither party is able to prevent or overcome.

**L. मध्यस्थता / Arbitration:**

If any dispute which may arise with respect to any term and condition or with respect to the interpretation of any term and condition of the Purchase Order/Work Order, which may be issued to the qualified and successful tenderer subsequently, the same shall be settled strictly in accordance with and in compliance of the Arbitration procedure which is mentioned descriptively in the Purchase Order/ Work Order.

**M. बोली अस्वीकृती के लिए मानदंड / Criteria for bid rejection:**

1 If technical bid found without the required undertaking (Annexure-II), EMD & Self Certificate (Annexure-IV)

2 If it found at a later date that any information given in the bids is Incorrect/false then the bid is liable to be disqualified/rejected.

3 Canvassing in any form will result to disqualification.

4 If “extra, as actual” etc. are mentioned against any of the price components in the bid the bid will be rejected.

5 If the firm quotes 'NIL' charges / consideration, the bid shall be treated as unresponsive and will not be considered.

6 If financial bid is found along with technical bid in pdf on CPP Portal, then the bid shall be rejected outrightly.

7 If any bidder submitting two/multiple bids for the same item of a tender, both/all the bids of that bidder will be outrightly rejected.

### Chapter-3 : Schedules of requirements

**N. टिप्पणी / Note:**

1 If the supplier/firm is manufacturer/authorized dealer/sole distributor/of the item, the certificate to this effect should be attached.

2 Please note that the firms must submit the compliance Statement in an organized and structured manner in respect of all the specifications as per Annexure-V with the supporting catalogue/leaflet of the firm. Annexure-I, II, & IV have to be submitted with Technical Bid and Annexure-III with financial bid.

3 It may also be noted that the tender bids received after the stipulated date & time or found incomplete and the tender bids containing false/incorrect information shall be summarily rejected. The Institute shall not entertain any communication in this regard, whatsoever.

4 The firm is required to give its Bank Account details with IFSC for the purpose of making e-payment.

5 The Indian Institute of Technology (IIT) Roorkee reserves the right to reject any quotation wholly or partly without assigning any reason.

6 The quantity shown in the tender can be increased or decreased to any extent depending upon the actual requirement.

7 The decision of the institute in all matters relating to eligibility, acceptance, rejection of the bid will be final and binding on the applicants.

8 IIT Roorkee discourages High Sea Sale purchase. All tenders with High Sea Sale will be rejected.

9 The firms should clearly mention (in the financial bid) the Harmonized System of Classification (HS code) defined by the Central Custom and Excise Board (Govt. of India)- for the item(s) involving import.

10 **If the bid opening date happens to be a holiday, then the bid will be opened at 3.30PM of the next working day followed by the holiday.**

11 If an agent submits bid on behalf of the Principal/OEM, the same agent can not submit a bid on behalf of another Principal/OEM in the same tender for the same item/product.

12 Indian Agent should submit the certificate from the Foreign Principal that they are the Authorized & Registered Indian Agent.

13 The quantity shown against the item is approximate and may vary as per demand of the Institute at the time of placing order. Also, the final requirement of each line item will be decided by the Institute at the time of placing the PO.

14 **Any corrigendum/addendum/errata in respect of the Tender Notice/ Enquiry letter shall be made available CPP Portal (<https://eprocure.gov.in/eprocure/app>). Hence prospective bidders are advised to visit the CPP Portal (<https://eprocure.gov.in/eprocure/app>) regularly.**

15 Bidders shall be responsible for the correctness of the information provided in the enquiry letter/tender document. If it is found at a later date that any information given in the bid is Incorrect/false then the bid is liable to be disqualified/rejected.

16 Before submission of bids, bidders should read the complete enquiry letter/tender document carefully and ensure that the bidders fulfill the eligibility criteria.

- 17 In case the item/product is under rate contract with DGS&D or GeM, then Bidder should certify that the rate quoted is not over and above (higher side) the rates under RC with the DGS&D or GeM.
- 18 **In Item Wise BoQ .xls sheet, In column M, leave blank for the item/service not provided/quoted by you. If quoted 0 then it will be considered that the item/service will be provided free of cost.**  
**Or**  
**In Item Rate BoQ .xls sheet, If quoted 0 for the items/services in column M then it will be considered that the items/services will be provided free of cost.**
- 19 **The bidder must quote the applicable GST appropriately in the price bid format provided i.e BoQ in .xls format. In case column of GST left blank, quoted 0, or inappropriate, then total quoted amount shows in column BB would be final inclusive of GST & other charges. The difference in GST amount payable – if any will be borne by the bidder. In this case, the basic unit prices will be suitably adjusted by the bidder – if required.**
- 20 **Only 'Class-I local supplier' and 'Class-II local supplier', as defined in the Public Procurement (Preference to Make in India), Order 2017 shall be eligible to bid in tender. For more details please refer: Order No.: P-45021/2/2017-PP (BE-II), DPIIT, Ministry of Commerce and Industry issued Dated: 16th Sept. 2020.**
- Self-certificate has to be provided in this regard that the item offered meets the local content requirement for 'Class-I local supplier' & 'Class-II local supplier'. (As per annexure-IV)**
- 21 **The Institute reserves the right to rectify any discrepancy of this advertisement, if found later on. In case of any inadvertent mistake in the process which may be detected at any stage even after the issue of purchase order, the Institute reserves the right to modify/withdraw/cancel any communication made to the bidder.**

**Chapter-4 : Specifications and allied Technical details**

**Plz. See Annexure- I, II, IV, V**

**Chapter-5 : Price Schedule (to be utilized by the bidders for quoting their prices online in .XLS format at CPPP Portal (<https://eprocure.gov.in/eprocure/app>))**

**Plz. See Annexure- III & Price Bid in .XLS Format**

**Chapter-6 : Contract Form**

**Purchase Order is used as Contract Form**

**Chapter-7 : Other Standard Forms, if any to be utilized by the purchaser and bidders**

**Tender Acceptance Letter**

\*उप कुलसचिव (सामग्री प्रबंधन)  
Deputy Registrar (MM)

**\*Document Digitally Signed on <https://eprocure.gov.in/eprocure/app>**

**For any Clarification regarding technical specifications of item Please Contact:**

**Head of the Department,  
Department of Chemistry,  
IIT Roorkee**

**Department of  
Chemistry,  
IIT Roorkee**

**Ph.-01332-28-5218  
Email: [head@cy.iitr.ac.in](mailto:head@cy.iitr.ac.in)**

तकनीकी बोली

विस्तृत विवरण अनुपालन रिपोर्ट के साथ प्रस्तुत किया जाना है (अपने आधिकारिक लेटर हेड पर विक्रेता/फर्म द्वारा प्रस्तुत करने के लिए )

Detailed specification cum compliance report to be submitted with the Technical Bid  
(to be submitted by the vendor/firm on its official letter head)

**(For Detailed Technical Specification Plz refer Annexure-V)**

Sr.No	Name & Required Specifications of Item(s)	Qty	UOM	Offered specifications of item(s) by the firm
1	Floor Mounted Powder X-ray Diffractometer	1	Set	
	Firm's Profile :			
1	Manufacturer/Authorized Agent/Distributor/Dealer/Supplier(valid certificate must be attached)			
2	E-mail I.D. & Telephone/Mobile No. Detail of Bank A/C			
Note:	1. "offered specifications", should be filled up properly in the format provided as per Annexure-I.			
	2. Volume & Weight of the item should also be mentioned.			
	3. The HSN / SAC Code of the item must be mentioned.			

Signature: \_\_\_\_\_

Name : \_\_\_\_\_

Designation : \_\_\_\_\_

Seal of the Organisation

उपबंध / UNDERTAKING

Ref. No. No. 2025260425/MM-13/IITR/2025-26/Floor Mounted Powder X-ra/CHEMISTRY/448  
Dated:- 12-Dec-25

I/We hereby certify that we or Our OEM are not from such a country which shares a land border with India or, if from such a country, We are registered with the Competent Authority. [Where applicable, evidence of valid registration by the Competent Authority is attached.]" (For details Please refer Order No.: F.No. 6/18/2019-PPD Dt.: 23rd July 2020)

Link of Order: [http://mm.iitr.ac.in/mmweb/public/forms/63374\\_GFR%20Rule%20144%20\(xi\).pdf](http://mm.iitr.ac.in/mmweb/public/forms/63374_GFR%20Rule%20144%20(xi).pdf)

That all the terms/conditions mentioned in the enquiry letter/tender against which the rates are being given are acceptable to the firm.

It is certified that the rate quoted is not over and above (higher side) the rates under RC with the DGS&D or GeM.

It is certified that the firm has never been black-listed from any Government Department. (State/Central Govt./ Autonomous/ PSU) in the last three years.

Signature \_\_\_\_\_

Name \_\_\_\_\_

Designation \_\_\_\_\_

Seal of the Firm/Agency

**Financial Bid**

(to be submitted by the vendor/firm on its official letter head)

Ref. No. No. 2025260425/MM-13/IITR/2025-26/Floor Mounted Powder X-ra/CHEMISTRY/448  
 Date: 12-Dec-25

Rates :

Sr. No	Name of item / Specifications	Qty	UOM	Rate	Amount (INR Only)
1	Floor Mounted Powder X-ray Diffractometer	1	Set	To be filled in .XLS sheet (attached)	
<b>Total</b>					
<b>Grand Total Rs.</b>					

1) Price of the items should be clearly mentioned if supplied Free of Cost.

## Terms &amp; Conditions:

- 1 Prices (quoted in Rs.): FOR Department of Chemistry,  
IIT Roorkee
- 2 **Payment:** **Should be clearly mentioned, As per clause-I of tender document.**
- 3 Validity:
- 4 Delivery period:
- 5 Other charges:
- 6 **Warranty:**
- 7 Rejection: Equipment if found not as per the required specification would be rejected.
- 8 Penalty Clause: As per clause-J of tender document.  
In case of the non-supply of the item/service after the acceptance of the work/supply order, EMD & PBG (if submitted) will be forfeited by the Institute. Also, the firm will be debarred from any future bidding process of IIT Roorkee for a period of a minimum two year.
- 9 PBG: In case the total order value of items becomes Rs.25.00 Lakhs or above, 5% of the total order value in the form of Bank Guarantee or FDR pledged in favor of "The Registrar, IIT Roorkee" is required at the time of installation of the equipment which should remain valid at least during the warranty period plus sixty days. No interest shall be claimed by the bidder on the PBG submitted.

It is certified that the rate quoted is not over and above (higher side) the rates under RC with the DGS&D or GeM.

Signature \_\_\_\_\_  
 Name \_\_\_\_\_  
 Designation \_\_\_\_\_  
 Seal Of the Firm/Organization \_\_\_\_\_

**Details of Institute :**

IEC CODE	0100000011	
AD CODE	0001069-2770325	S.B.I., IIT ROORKEE
	0303974-2770324	P.N.B. IIT ROORKEE

**(to be submitted by the vendor/firm on its official letter head)**

**Self-Certificate for Local Content**

**Tender No.:** No. 2025260425/MM-13/IITR/2025-26/Floor Mounted Powder  
X-ra/CHEMISTRY/448

**12-Dec-25**

We hereby certify that the items quoted by us against above mentioned tender no. has the local content as per below:

**Local Content (in %):.....**

**Local Supplier Class:.....**

The details of the Make in India items/parts used in the quoted products is/are as under:

- 1
- 2
- 3

The details of the location(s) at which the local value addition made/manufactured is/are as under:

- 1
- 2
- 3

We also understand, false declarations will be in breach of the code of integrity under rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for up to two years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under law.

**Signature:**

**Name:**

**Designation:**

**Seal Of the Firm/Organization:**

## Instructions for Online Bid Submission

Instructions to the Bidders to submit the bids online through the Central Public Procurement Portal for e-Procurement at <https://eprocure.gov.in/eprocure/app>. Also, any modification on CPP portal in reference to below will be applicable. Therefore, for more details and update please refer CPP Portal (<https://eprocure.gov.in/eprocure/app>).

- 1 Possession of valid Digital Signature Certificate (DSC) and enrollment/registration of the contractors/bidders on the e-Procurement/e-tender portal are prerequisite for etendering.
- 2 Bidder should register for the enrollment in the e-Procurement site using the "Online Bidder Enrollment" option available on the home page. Portal enrollment is generally free of charge. During enrollment/registration, the bidders should provide only valid and true information including valid email id. All the correspondence shall be made directly with the contractors/bidders through email id as registered.
- 3 Bidder need to login to the site through their user ID/ password chosen during enrollment/registration.
- 4 Then the Digital Signature Certificate (Class III Certificates with signing key usage) issued by SIFY/nCode/eMudra etc. recognized by Controller of Certifying Authorities (CCA) India on eToken/SmartCard, should be registered.
- 5 The registered DSC only should be used by the bidder in the transactions and should ensure safety of the same.
- 6 Contractor/Bidder may go through the tenders published on the site and download the tender documents/ schedules for the tenders.
- 7 After downloading/getting the tender document/schedules, the Bidder should go through them carefully and then submit the documents as required, otherwise bid will be rejected.
- 8 Any clarifications may be sought online through the tender site, through the contact details or during pre-bid meeting if any. Bidder should take into account the corrigendum if any published before submitting the bids online.
- 9 Bidder may log in to the site through the secured login by the user id/ password chosen during enrolment/registration and then by submitting the password of the e-Token/Smartcard to access DSC.
- 10 Bidder may select the tender in which he/she is interested in by using the search option and then move it to the 'my tenders' folder.
- 11 From my tender folder, he may select the tender to view all the details uploaded there.
- 12 It shall be deemed that the bidder has read and understood all the terms and conditions before submitting the offer. Bidder should go through the tender schedules carefully and upload the documents as asked; otherwise, the incomplete bid shall stand rejected.
- 13 Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document/ schedule and ordinarily it shall be in PDF/xls/rar/jpg/dwf formats. If there is more than one document, all may be clubbed together and provided in the requested format. Bidders Bid documents may be scanned with 100 dpi with black and white option. It is advisable that each document to be uploaded through online for the tenders should be less than 2 MB. If any document is more than 2MB, it can be reduced through zip/rar and the same if permitted may be uploaded. The file size being less than 1 MB the transaction uploading time will be very fast.
- 14 The Bidders can update well in advance, the documents such as certificates, annual report details etc., under "My Space option" and these can be selected as per tender requirements and then send along with bid documents during bid submission. This will facilitate the bid submission process faster by reducing upload time of bids.
- 15 Bidder should submit the Tender Fee/ EMD as specified in the tender. Scanned copy of the same should be uploaded as part of the offer. (if applicable)
- 16 While submitting the bids online, the bidder shall read the terms and conditions and may accept the same to proceed further to submit the bid packets.
- 17 The bidder has to select the payment option as offline to pay the Tender FEE/ EMD as applicable and enter details of the same. (if applicable)
- 18 The details of the DD/any other accepted instrument, physically delivered, should tally with the details available in the scanned copy and the data entered during bid submission time, otherwise submitted bid shall not be acceptable or liable for rejection. (If applicable)
- 19 The bidder has to digitally sign and upload the required bid documents one by one as indicated. Every act of using DSC for downloading the bids and uploading their offers shall be deemed to be a confirmation that they have read, understood and agreed with all clauses of the bid document including General conditions of contract without any exception.

- 20 The bidder has to upload the relevant files required as indicated in the cover content. In case of any irrelevant files, the bid may be rejected.
- 21 If the price bid format is provided in a spread sheet file like BoQ\_XXXX.xls, the rates offered should be entered in the allotted space only and uploaded after filling the relevant columns. The Priced-bid/BOQ template shall not be modified / replaced by the bidder; else the bid submitted is liable to be rejected for the tender.
- 22 The bidders are advised to submit the bids through online e-tendering system to the Tender Inviting Authority (TIA) well before the bid submission due date and time (as per Server System Clock). The TIA shall not be held responsible for any delay or the difficulties faced during the submission of bids online by the bidders.
- 23 After the bid submission (i.e. after Clicking "Freeze Bid Submission" in the portal), the acknowledgement number indicated by the system should be printed by the bidder and kept as a record of evidence for online submission of bid for the particular tender and also be used as entry pass to participate in the bid opening.
- 24 The time settings fixed in the server side and displayed at the top of the tender site, shall remain valid for all actions of requesting, bid submission, bid opening etc., in the e- Tender system. The bidders should follow such time during bid submission.
- 25 All the data being entered by the bidders would be encrypted using Public Key Infrastructure (PKI) encryption techniques to ensure the secrecy of the data. The data entered is not retrievable by unauthorized persons during the bid submission and until the time of bid opening by any person.
- 26 Any bid document that is uploaded to the server is subjected to symmetric encryption using a system generated symmetric key. Further this key is subjected to asymmetric encryption using buyers/bid openers' public keys. Overall, the uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- 27 The confidentiality of the bids is maintained with the use of Secured Socket Layer (SSL) 128 bit encryption technology. Data storage encryption of sensitive fields is done.
- 28 The bidder should logout of the tendering system using the normal logout option available at the top right hand corner and not by selecting the (X) exit option in the browser.
- 29 For any queries regarding e-Tendering process, the bidders may contact at address as provided in the tender document. Parallely for any further queries, the bidders are advised to contact over phone: **0120-4001005** or send an e-mail to – [cppp-nic@nic.in](mailto:cppp-nic@nic.in).

### **Prequalification**

- 1 An undertaking by the firm that it has never been black-listed in the last three years must be attached along with the Bid, failing which the Bid shall be rejected. (As per Annexure-II)
- 2 True copy of Permanent Account Number.
- 3 Details of GST along with a copy of certificate to be attached.
- 4 Submission of samples if required, for all items indicated in the NIT. The make of items proposed to be supplied should be indicated and submitted along with the techno commercial unpriced bid without indicating the pricing components.
- 5 Willingness to execute all orders which are placed to meet emergency requirement on priority basis. The Bidder shall note that standards for workmanship, material and equipment, and references to brand names designated by the Purchaser in the tender document are intended to be descriptive only and not restrictive. The Bidder may substitute alternative standards, brand names and/or catalogue numbers in his bid, provided that it demonstrates to the Purchaser's satisfaction that the substitutions ensure substantial equivalence to those designated in the Technical Specifications.

**Tender Acceptance Letter**  
(To be given on Company Letter Head)

---

---

---

---

---

**Date:**

**Sub: Acceptance of Terms & Conditions of Tender.**

**Tender No. 2025260425/MM-13/IITR/2025-26/Floor Mounted Powder X-ra/CHEMISTRY/448**

Tender description:

---

---

---

---

---

Dear Sir,

- 1 I/We have downloaded/ obtained the tender document(s) for the above mentioned 'Tender No.' from the web site(s) namely:  
  
as per your advertisement, given in the above mentioned website(s).
- 2 I/We hereby certify that I/We have read the entire terms and conditions of the tender documents from Page No. \_\_\_\_\_ to \_\_\_\_\_ (including all documents like section(s), schedules(s) etc.), which form part of the contract agreement and I/we shall abide hereby by the terms/conditions/ clauses contained therein.
- 3 The corrigendum(s) issued from time to time by your department/ organisation too have also been taken into consideration, while submitting this acceptance letter.
- 4 I/We hereby unconditionally accept the tender conditions of above mentioned tender document(s)/ corrigendum(s) in its totality/entirety.
- 5 In case any provisions of this tender are found violated, then your department/organisation shall without prejudice to any other right or remedy be at liberty to reject this tender/bid including the forfeiture of the full said earnest money deposit absolutely.

**Yours Faithfully,**

**(Signature of the Bidder, with Official Seal)**

**DECLARATION**  
**(to be submitted by the vendor/firm on its official letter head)**

**To,  
Registrar,  
IIT Roorkee,  
Roorkee,**

Subject: Undertaking in lieu of Bid Security/Earnest Money Deposit (EMD)

Tender Ref no: No. 2025260425/MM-13/IITR/2025-26/Floor Mounted Powder X-ra/CHEMISTRY/448

12-Dec-25

Dear Sir,

We, the undersigned, offer to Supply the .....  
..... as per tender at IIT Roorkee, in response to your above refered tender no.

We are hereby submitting our proposal/Bid/Quote for the same, which includes Technical bid and the Financial Bid. As a part of eligibility requirement stipulated in above refered tender document, we hereby submit a declaration in lieu of Earnest Money Deposit (EMD) as given below:

1. Our bid shall remain valid for 90-days from the date of financial opening and that we will not withdraw or modify our bid during the validity period,
2. In case, we are declared as successful bidder and an order is placed on us, we will submit the acceptance in writing within 7 days of placement of order on us.
3. In case, we are declared as successful bidder and an order is placed on us, we undertake, to submit a Performance Bank Guarrantee(PBG) as mentioned in above refered tender document or as per terms stipulated in the Purchase Order.
4. In case of failure on our part to comply with any of the above said requirements, we are aware that we shall be declared as un-eligible for said tender and /or debarred from any future bidding process of IIT Roorkee for a period of minimum one year.
5. The undersigned is authorized to sign this undertaking.

**Signature:**

**Name:**

**Designation:**

**Seal of the Firm/Organization:**

<b>Technical Specifications for Floor Mounted Powder X-ray Diffractometer</b>	
<b>Objective:</b> Supplying, Installation, and Commissioning of a 3kW, high resolution X-Ray Diffractometer (XRD) required for powder diffraction and phase analysis of powders, polycrystalline thin film (GIXRD), SAXS for particle and pore size analysis and a high-temperature attachment. The system should be fully computer-controlled with hardware and software for the measurements. It should be capable of working in reflection as well as in transmission mode. System should be supported by manufacturers for the supply of spares for at least 10 years after installation. The instrument shall comprise the following detailed specifications:	
<b>1</b>	<b>Essential Analytical Capability:</b>
	The system must be capable of sensing and recognizing all the components in the beam path, including X-Ray tube, optics in both primary and secondary sides, all types of sample stages and detectors to avoid any mistakes while configuring the system for different types of analytical experiments. For fast changeover of modes of operation, the XRD system must have auto-alignment facility for entire optical path and stages. Visibility of the instrument inside a minimum of 75% of the frame cabinet.
<b>2</b>	<b>X-Ray Generator</b>
	<ul style="list-style-type: none"> <li>i) Output power: 3.0 kW or higher.</li> <li>ii) Maximum Voltage: 50 kV or higher with an increment of 1.0 kV</li> <li>iii) Maximum Anode Current: 60 mA or higher with an increment of 1.0 mA</li> <li>iv) Control: Fully automated through software</li> </ul>
<b>3</b>	<b>Goniometer</b>
	<ul style="list-style-type: none"> <li>i. Type: Higher resolution Theta/Theta (independent or coupled), Vertical</li> <li>ii. The Goniometer should have independent <math>\theta</math> and <math>2\theta</math> drives for accurate positioning, and with a design to have zero backlash.</li> <li>iii. Two theta range: from <math>-5^\circ</math> to <math>160^\circ</math> or higher for reflection geometry, with capability of vertical transmission mode while keeping the sample stage in horizontal position to avoid any sample spill in both reflection and transmission geometries.</li> <li>iv. Scan Speed Between 0.001 to 1.25 deg/sec or better</li> <li>v. Minimum step size: <math>0.0001^\circ</math> or better</li> <li>vi. Goniometer radius: 240 mm or more. Suitable hardware and software to be included for the goniometer radius change facility.</li> </ul>
<b>4</b>	<b>X-Ray Tube</b>
	<ul style="list-style-type: none"> <li>i) Target: Copper (Cu), Optional Co source to be quoted in the tender as an optional accessory.</li> <li>ii) Focus: Long fine focus (LFF)</li> <li>iii) Insulation: Ceramic insulation.</li> <li>iv) Max. output power: <math>\geq 2.0</math> kW</li> <li>v) Suitable Beta filter for X-ray tubes to be included</li> <li>vi) X-ray tube must have an auto component recognition facility to indicate anode target type, its operation time and its exchange from one source type to another type and changeover from line to point focus type should be with an auto alignment facility.</li> <li>vii) Tube height and tube focus for different types of optics must have an auto-alignment facility.</li> <li>viii) Provision for change from line focus to point focus without any disconnection of power and water supply</li> </ul>
<b>5</b>	<b>Sample stage and Sample holder</b>

	<ul style="list-style-type: none"> <li>i) Sample stage for powder sample suitable for both reflection and transmission geometries with facility to keep powder sample always in horizontal position with a set of reflection and transmission holders.</li> <li>ii) Sample stage should have minimum 7 mm movement along z-axis for sample alignment.</li> <li>iii) System must have both automatic and manual sample alignment options for bulk, thin film samples alignment.</li> <li>iv) Standard sample stage for powder diffraction which can allow the powder analysis.</li> <li>v) 20 numbers of Glass/PMMA/SS/Alumina sample holder with 0.2-1.0 mm depression</li> <li>vi) 1 number of Si crystal low background sample holders without depression</li> <li>vii) 1 number of Si crystal low background sample holder with 0.5-1.0 mm depression</li> <li>viii) 1 number of air tight sample holder</li> <li>ix) 1 number of sample holder (with variable spacers or similar functionality) to accommodate thin film sample upto the height of 21 mm.</li> <li>x) 1 number of sample holder for transmission geometry</li> <li>xi) Appropriate optics or mechanics to reduce the background at lower 2<math>\theta</math> range.</li> <li>xii) SAXS-Appropriate Slit, optics, sample holder, path and software for SAXS to analyze particle size and pore size distribution. Necessary optics and accessories to measure the angle in the range 0.2 to 10<math>^\circ</math> applicable for SAXS application with (i) monochromatic (ii) high intensity collimated beam.</li> </ul>
<b>6</b>	<b>Optics</b>
	Required number of all optical components i.e. soller slits, divergent slits, anti-scatter slits, parallel slits, mask and beta filters and other necessary components for full operation should be provided. Optical component settings must be fully automatically recognized, and aligned by XRD control software with the latest technologies for error-free analysis.
<b>7</b>	<b>Detector</b>
	<ul style="list-style-type: none"> <li>i) 0D, Scanning and fixed 1D mode should be possible.</li> <li>ii) Global count rate <math>\geq 10^8</math> CPS.</li> <li>iii) Pixel/strip size should be 75 micron or less.</li> <li>iv) Number of pixels/sensors/strips: 190 <math>\times</math> 190 or higher.</li> <li>v) Number of channels: 190 or higher</li> <li>vi) Detector should work in continuous, fast scan, slow-scan, static modes.</li> <li>vii) Detector should be supplied with the capability to suppress secondary fluorescence</li> </ul>
<b>8</b>	<b>High temperature stage</b>
	High temperature stage with strip heating from room temperature to 1400 $^\circ$ C using direct heating with all necessary accessories, including a temperature controller, sample holder, vacuum pump, gas cylinder (Ar, N $_2$ ) with regulator and all other mandatory accessories
<b>9</b>	<b>Standards</b>
	<ul style="list-style-type: none"> <li>i) Fluorescent disc for positioning the X-ray beam to be supplied with the system.</li> <li>ii) Necessary standard sample for checking the accuracy of the peak position over the entire goniometer angular range.</li> </ul>
<b>10</b>	<b>Software</b>

	<p>i) PC based software packages to control all instrument parameters and should provide best algorithms for Solving analytical tasks. Should enable both Qualitative and Semi quantitative analysis. Easy simulation and automatic refinement / smoothing of measured data. Peak search, integrated intensity, K alpha 1 and K alpha 2 separation.</p> <p>ii) Search match software with full-pattern-approach to phase identification with an integrated, quantitative phase analysis module to be offered. Data analysis software should be with minimum 5 users licenses for use by multiple scientists and researchers.</p> <p>iii) ICDD PDF 5+ Database with 5-year license extension to be offered for organic, inorganic and Rietveld analysis.</p> <p>iv) Rietveld-based standard less quantification, refinement package with ab initio structure analysis, profile fitting, lattice parameter refinement, indexing, and crystallite size determination should be offered.</p>
11	<b>Computer:</b> A desktop computer (HP/Lenevo/Dell) of suitable configuration compatible with the instrument should be provided with the XRD instrument, with an OLED monitor of size at least 24 inches. Another desktop computer with i9 core, 1 TB hard disk (SSD), 16 GB ram and 32" inch OLED monitor with the aforementioned software to be supplied as a data processing station.
12	<b>Safety:</b> Provision for detection of abnormal cooling water flow, pressure and temperature, abnormal generator overload and shutter malfunction detection, and emergency shutdown. The XRD instrument should be compliant as per International Safety Standards and Regulation norms like Eurotom, pertaining to mechanical, electrical and X-ray Radiation safety and other hazards. The supplier must have the necessary TYPE approval from AERB for the supply of XRD and the necessary approval document to be submitted along with the offer.
13	<b>Water Chiller for X-ray source:</b> A compatible external water chiller from Polyscience/Klosco/Werner Finley/Hexon should be offered separately with the capability to work at an ambient temperature of 40 degrees C or more. External filter/traps should be provided to minimize the dust settlements from the chiller to the target area.
14	<b>UPS:</b> Minimum 20 KVA or more <b>IGBT-based</b> online UPS with built-in isolation transformer. UPS must be able to handle the XRD system, computer, and external water chiller pump for a minimum of 60 minutes of uninterrupted XRD data collection during a main power failure. UPS should be quoted from one of the following brands: Leibert/Vertiv/APC/Microtek
15	<b>Other mandatory requirements</b> Pre-installation requirements must be provided upon as and when required and the supplier to arrange a site inspection for installation planning. The vendor should provide all required accessories, like tools and tackle, for the complete installation and commissioning of the instrument. Buyer to arrange necessary power connection, wiring, earthing, air conditioning, flooring and other furniture in the lab for smooth installation of equipment. <ul style="list-style-type: none"> <li>• The complete set of documentations, manuals and software manuals should be provided both in Softcopy and hard copy</li> <li>• Spare parts and consumable necessary to run the instrument for at least three years</li> <li>• <b>Certification:</b> The supplier will certify that the whole package of equipment will be capable of doing (i) SAXS and, (ii) powder diffraction and analysis in all above mentioned working conditions with respective data acquisition and storage, in an integrated manner, without need for any other accessories.</li> </ul>

25-11-2025  
 25-11-2025  
 Page 16 of 17

	<ul style="list-style-type: none"> <li>• <b>Compliance statement:</b> Point wise according to our document</li> <li>• <b>User List:</b> Provide user list with contact details (address and phone) of the quoted instrument (model) for last 3 years</li> </ul>
16	<p><b>Application and Operation Training:</b>  Supplier must provide complete operation, do's/don'ts and application training for all aforesaid applications for a minimum of 3 days. All costs towards training, including lodging, boarding and travel, have to be borne by the XRD supplier.</p>
17	<p><b>Warranty:</b>  The quoted XRD System with all accessories must be provided with a complete &amp; comprehensive warranty for 1 year, including X-Ray Generator, X-ray tube, electronics, motors, and detector. During the warranty period, if any part is required for replacement, then the same is to be arranged on a DDP basis without any cost to the buyer. The instrument warranty will be extended for any period of downtime due to the malfunctioning of the instrument during the warranty period.  Post standard one year warranty, an additional comprehensive AMC of 4 years is to be provided in the tender.</p>

Redner

P  
25/11/25

Ⓜ

Redner

For  
25/11/25

25-11-2025